



A Scan SW-1000 SW-1000P

A scan probe: 10MHz import small size probe,
built-in luminotron

Measuring range: 15mm-40mm

Measurement precision: ± 0.05 mm; with macula
lutea trace function

Measurement: Anterior chamber depth, lens
thickness, vitreous body length, total length and
average Method of measurement: immersion and contact

Eye mode: Phakic/ Aphakic/Dense/ various IOL

IOL formula: SRK-II、SRK-T、BINKHORST-II、HOLLADAY、HOFFER-Q、
HAIGIS

Enter the name & ID; easy to check archive

Storage: 10 cases, 5 readings each case

Output: A scan waveform and IOL calculation sheet

P scan probe: 20MHZ, angle of 45 degrees makes easier operation

Resolution: 5um

Measuring range: 150um~1500um

Display: SINGLE mode and MAP mode

Can display ultrasound waveform when measuring

Each group is the average of 20 measurements

Switch between IOP measured value and actual value

Can input name, ID and operator's name

Others:

Large color liquid-crystal screen

Touch screen input, easy operation

Curve freezing: Manual/Auto mode,
controlled by pedal

Built-in speed thermal printer



SW-1000-GG-00



SW-1000 SW-1000P



A Scan

SUOER A Scan SW-1000: with macular recognition function, can accurately measure axial
length, anterior chamber depth, lens thickness, as well as intra-ocular lens calculation.

SUOER Pachymeter SW-1000P: with the central area and the peripheral areas gain automatic
compensation capability, accurate measurement of central and peripheral corneal thickness,
is widely used in the preoperative examination and postoperative effect evaluation of refractive
surgery.

Company Name: Tianjin Suwei Electronic Technology Co., Ltd.
Address: Room 2-201,2#Building,No.6,Zhuyuan Road,Huayuan
Industrial Zone,Tianjin,China
Tel: +86-22-8371 2745
Fax: +86-22-8371 2645
E-mail:sales@suwei.com.cn

Company Name: Suwei Technology Service (U.S.), Inc
Address: 415E. Golf Road,Suite 101-120 Arlington Heights,
IL,60005, US
Tel: +1-847-201-4523
E-mail:sales@suwei.com.cn

SUOER